

VERIFICATION OF CALIBRATION REPORT

OCT 05 2009

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130136
A	Name <u>Nita Bolz</u>	ID# <u>7384</u>	Date <u>10/2/09</u>
	Agency <u>AST Bethel</u>	Phone # <u>269-0599</u>	
	Instrument Location <u>AST Bethel</u>		
B	Alco S/N <u>77675</u>	Target Value <u>0.086</u>	High Pressure <u>1200</u>
	Alco Test Values <u>0.91</u>	<u>0.92</u>	
	1 st Alco	2 nd Alco	
	Signature <u>Nita Bolz</u>	(OVER)	

8MB 10/13/09

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 22nd day of Oct., 2009.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130136 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136

OCTOBER 02, 2009

OPERATOR'S NAME:

BOLZ/WITG/J

OPERATOR'S NUMBER: 7034

SUBJECT'S LAST NAME:

NO

SUBJECT'S FIRST NAME:

NO

DATE: 01/02/09

DEPT/AGENCY: SE7

LAB/REPORT: NO

TEST TYPE: 0

ALCOHOL TEST: 0.00 ✓

ALCOHOL TEST: 0.00 ✓

BREATH ANALYSIS

ALCOHOL ADJUSTED P/B 0.00 ✓

ALCOHOL TEST: 0.00 ✓

ALCOHOL TEST: 0.00 ✓

INTERNAL STANDARD: 0.00 ✓

ALCOHOL TEST: 0.00 ✓

BLANK TEST: 0.00 ✓

INTERNAL STANDARD: 0.00 ✓

BLANK TEST: 0.00 ✓

ALCOHOL TEST: 0.00 ✓

BLANK TEST: 0.00 ✓

DIAG

CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130136 ✓

OCTOBER 02, 2009

TIME: 10:35

DIAGNOSTIC CHECK

COMPUTER: OK ✓

PROGRAM: OK ✓

SOFTWARE DATE: 01/08/09

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.99

FLOW TEST: 0.00

VALVE

VALVE: 0.00

DETECTOR: OK ✓

FILTERS: OK ✓

EXHAUST STANDARD: OK ✓

CALIBRATION: OK ✓

PRINTER TEST

ALCOHOL TEST: 0.00 ✓

INTERNAL STANDARD: 0.00 ✓

BLANK TEST: 0.00 ✓